



HISTORY

Since 1988, we have specialized in the development, manufacturing, and distribution of protective infection control solutions.

At Cranberry, we stress utmost importance on providing products of superior quality. Therefore, we constantly strive to develop products with the highest protection, comfort, and strength. After thriving in this industry for so many years, we truly understand your needs and demands as a professional. Be assured that our dynamic team is always moving forward, researching, and seeking to provide you with only the best.

We have established successful partnerships in many countries. Even so, we are excited to expand our networks and distribution further so that Cranberry gloves are made available in every country.

As a professional, you are passionate about the health and comfort of your clients. At Cranberry, we are passionate about yours.

FIRST TOUCH



In an average manufacturing process, gloves and masks may come in human skin contact up to 8 times. With ultimate hygiene in mind, Cranberry products are First Touch® manufactured, examined, and packaged with zero direct skin contact exposure. Don't just put on any gloves and masks, look for our First Touch® logo and be assured that you are doing the best you can to protect yourself and your patients.

CERTIFICATIONS




Quality Management Systems



ISO 13485 & ISO 9001



TGA
AUSTRALIA





DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 19 1999

Mr. [REDACTED]
Director
Cranberry (M) [REDACTED]
[REDACTED]

Re: [REDACTED]
Trade Name: Nitrile Powder-Free Examination Gloves
Regulatory Class: I
Product Code: LZA
Dated: January 15, 1999
Received: January 19, 1999

Dear Mr. [REDACTED]

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Good Manufacturing Practice for Medical Devices: General (GMP) regulation (21 CFR Part 820) and that, through periodic GMP inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of


Page 2 - Mr. [REDACTED]

the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4692. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,



Timothy A. Ulatowski
Director

Division of Dental, Infection Control,
and General Hospital Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



CRANBERRY (M) [REDACTED]

Applicant : CRANBERRY (M) SDN. BHD.
[REDACTED]

510(k) Number (if known): [REDACTED]

Device Name: Nitrile Powder Free Examination Glove - Cranberry

Indications For Use:

This product is a patient examination glove. It is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Chin S. Lin

(Division Sign-Off)
Division of Dental, Infection Control,
and General Hospital Devices

510(k) Number [REDACTED]

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter-Use X

Test Report No. 7191
dated 23 Mar 2018



PSB Singapore

Note: This report is issued subject to the Testing and Certification Regulations of the TUV SUD Group and the General Terms and Conditions of Business of TUV SUD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Powder Free Nitrile Examination Gloves submitted by
 [redacted] Co., Ltd. on 25 Jan 2018 and 09 Mar 2018.

Choose certainty.
 Add value.

TESTED FOR:

[redacted]
 [redacted]
 [redacted] Shandong,
 China.

TEST DATE:

26 Jan 2018 to 06 Feb 2018 and 22 Mar 2018

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Reference No.	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Powder Free Nitrile Examination Gloves	Blue	BS0002	12150511	XS	100	[redacted] Ltd.
				03060511		69	
				12150521	S	100	
				12080311	M	100	
				12090411	L	100	
				12070611	XL	407	

Lot size as specified by client: 200,000 pieces

METHOD OF TEST:

- EN 455-1:2000 Medical gloves for single use
 Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
 Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
 Part 3: Requirements and testing for biological evaluation



Laboratory:
 TUV SUD PSB Pte. Ltd.
 No.1 Science Park Drive
 Singapore 118271

Phone: +65 6895 1333
 Fax: +65 6776 8670
 E-mail: enquiry@tuv-sud-psb.sg
 www.tuv-sud-psb.sg
 Co. Reg: 100009670

Regional Head Office:
 TUV SUD Asia Pacific Pte. Ltd.
 1 Science Park Drive, #02-01
 Singapore 118271

Test Report No. 719
dated 23 Mar 2018



RESULTS:

Sample: Powder Free Nitrile Examination Gloves, BS0002

Table 1: Results for EN 455-1:2000

Clause	Tests	Size	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	XS	Shall not leak	10	315	1	Passed
		S		10	315	1	Passed
		M		10	315	2	Passed
		L		10	315	6	Passed
		XL		10	315	4	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-6

Clause	Tests	Size	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	XS	≥ 240	13	250	Passed
		S		13	250	Passed
		M		13	255	Passed
		L		13	250	Passed
		XL		13	248	Passed
	b) Width (mm)	XS	≤ 80	13	73	Passed
		S	80 ± 10	13	85	Passed
		M	95 ± 10	13	96	Passed
		L	110 ± 10	13	106	Passed
		XL	≥ 110	13	115	Passed
5	a) Force at break (N)	XS	For nitrile examination gloves: ≥ 6.0	13	6.4	Passed
		S		13	8.6	Passed
		M		13	6.1	Passed
		L		13	6.1	Passed
		XL		13	6.6	Passed
	b) Force at break after challenge testing (N)	XS	For nitrile examination gloves: ≥ 6.0	13	7.0	Passed
		S		13	9.0	Passed
		M		13	7.3	Passed
		L		13	6.6	Passed
		XL		13	7.1	Passed

Table 3: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed

Test Report No. 7191
dated 23 Mar 2018



RESULTS (cont'd):

Sample: Powder Free Nitrile Examination Gloves, BS0002

Table 4: Results for EN 455-3:2015 Clauses 4.2-4.5

Clause	Tests	Requirements	Results / Remarks	Inferred results	
4.2	Chemicals	Gloves shall not be dressed with talcum powder (magnesium silicate).	Glove is powder-free glove, based on client's declaration letter version 2018001	NA	
		Other chemicals	Manufacturer shall disclose upon request a list of chemical ingredients	NA	
4.3 5.1	Endotoxins	< 20 EU/pair for gloves labelled with 'low endotoxin content'.	Not labelled with 'low endotoxin content'	NA	
4.4 5.2	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	XS	0.61 mg per glove	Passed
			S	0.63 mg per glove	Passed
			M	0.97 mg per glove	Passed
			L	0.86 mg per glove	Passed
			XL	0.48 mg per glove	Passed
4.5 5.3	Proteins, leachable	The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.	Non-natural rubber latex glove	NA	

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex: The labelling shall include the following or equivalent warning statement together with the symbol: '(Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses';	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free;	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: 'CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions';	NA
		d) for any medical glove containing natural rubber latex the product labelling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unjustified indication of the presence of allergens;	NA
		e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA
Inferred results			Passed



REMARKS:

1. Freedom from holes test for XS, S, M and L sizes were tested in manufacturer's site, witnessed by TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch on 22 Mar 2018.
2. For size XS, results for EN 455-2:2015 Clause 4 Dimensions is based on lot no. 03060511, while the rest of the results are based on lot no. 12150511.
3. Labelling requirements are assessed based on submitted packaging artwork together with client's declaration letter version number 2018003.
4. NA: Not applicable for the submitted sample.


Shareen Chan
Engineer


Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo : Powder Free Nitrile Examination Gloves, BS0002



PPE REGULATION (EU) 2016/425 MODULE C2 CERTIFICATE

Issued to:

Blue Sail Medical Co Ltd
Qilu Chemical Industrial Park
No 21 Dinglian Road
Zibo
Shandong
China

This is to certify that the following products tested under SATRA reports referenced: CHM0291439/1944/JH & STE0289547 have been found to satisfy the requirement of PPE Regulation (EU) 2016/425 Module C2 EU quality control system for the final product for and on behalf of SATRA Technology Europe Limited

EU TYPE EXAMINATION CERTIFICATE NUMBER	PRODUCT GROUP REFERENCE	PRODUCT TYPE	CLASSIFICATION
2777/11521-01/E00-00	BS01020X	Disposable medical Nitrile examination glove	EN ISO 374-1:016

Dated: 14th November 2019

This certificate is
valid until: November 2020

Signed By (Alan Weston)

For and on behalf of SATRA Technology
Europe Limited



The issuance of this certificate is subject to the company maintaining its manufacturing and quality system to the required standard.

SATRA Technology Europe Limited, Bracktown Business Park Clonsilla Dublin 15 D15 YN2P, Republic of Ireland.
(Notified Body number 2777)

Tel: +353 (0) 1 437 2484 Web: www.satrathegroup.com



CRANBERRY (M) SDN. BHD. (104994 - W)

Applicant : CRANBERRY (M) [REDACTED]

510(k) Number (if known): K [REDACTED]

Device Name: Nitrile Powder Free Examination Glove - Cranberry

Indications For Use:

This product is a patient examination glove. It is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Chin S. Lin
(Division Sign-Off)
Division of Dental, Infection Control,
and General Hospital Devices
510(k) N K98****

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter-Use X

C

Contour[®]

Nitrile Powder Free
Examination Gloves

Contour is body heat activated to contour ergonomically for reduced hand fatigue during periods of extended wear. In addition, the gloves provide latex-like comfort and superior donning efficiency. Available in blue, the gloves are packed 100 pieces per box.

