



Hartalega

gloveen COATS™

Colloidal Oatmeal System



XS-L: 100 Gloves
XL: 90 Gloves
By Weight


Made in Malaysia by Hartalega Sdn Bhd.

FEATURES

- Fingertip textured
- Powder Free
- Not made with natural rubber latex
- Chemo drugs tested
- Lab chemical tested
- Ambidextrous
- Standard cuff
- Dawn blue colour



HOW COATS® GLOVE WORKS



1 Barrier

COATS creates a physical barrier between glove and skin, hence preventing direct irritation from glove film.

2 Absorption

COATS absorbs water, urea and salt excretions that cause irritation.

3 Moisturisation

COATS penetrates quickly into skin to keep it moisturised.

COATS® Colloidal Oatmeal Coated Nitrile Powder Free 2.5 Mil

ASTM D3578

Physical Dimensions		
Glove Length (mm)	≥ 230	
Palm Thickness (mm)	0.07 ± 0.02	
Finger Thickness (mm)	0.09 ± 0.02	
Physical Properties		
Test	Before Aging	After Aging
Tensile strength (MPa)	≥ 18.0	≥ 16.0
Elongation (%)	≥ 500	≥ 400

EN 455

Physical Dimensions		
Median glove length (mm)	≥ 240	
Median palm thickness (mm)	0.07 ± 0.02	
Median finger thickness (mm)	0.09 ± 0.02	
Physical Properties		
Test	Before Aging	After Aging
Median Force at break (N)	≥ 6	≥ 6



Regulatory Compliance

FDA 510(k), MDD 93/42/EEC, REACH, ROHS Directive 2002/95/EC, EC 10/2011, EC 1935/2004, PPE 89/686/EEC

Standards

ASTM D6319, ASTM 6978, EN455 part 1, 2, 3 & 4, EN 1186, EN 13130, CEN/TS 14234, EN 420, EN 374 part 1, 2 & 3

Classification

Class I (FDA), Class I (MDD 93/42/EEC), Category 3 (BfR XXI), Category III (PPE 89/686/EEC)

Patent

7,691,436; 7,718,240; 7,740,622; 8,075,965; 8,458,818

Application Settings

Low risk - medical, dental, procedures, chemotherapy drugs, pathology lab and food handling. Coated with FDA recognised skin protectant. Clinically proven to help protect and moisturise your skin from dry and irritated skin from prolonged glove use and hand wash.

Colour

Dawn blue, white

CERTIFICAT



CERTIFICADO



СЕРТИФИКАТ



認證證書



CERTIFICATE



ZERTIFIKAT



Management Service

CERTIFICATE

The Certification Body
of TÜV SÜD Management Service GmbH

certifies that



Hartalega Sdn. Bhd.

C-G-9, Jalan Dataran SD1, Dataran SD PJU9, Bandar Sri Damansara,
52200 Kuala Lumpur
Malaysia

Scope of application:

**Distribution of Natural Latex and Nitrile Powdered
and Powder-Free Non-Sterile and Sterile Examination Gloves,
Sterile Surgical Gloves and Industrial Gloves**

Hartalega Sdn. Bhd.

No.7, Kawasan Perusahaan Suria, Bestari Jaya
45600 Selangor Darul Ehsan
Malaysia

Scope of application:

**Design and Development, Production and Distribution
of Natural Latex and Nitrile Powdered and Powder-Free Non-Sterile
and Sterile Examination Gloves, Sterile Surgical Gloves
and Industrial Gloves**

has established and applies
a Quality Management System.

An audit was performed, Order No. **721423945**.

Proof has been furnished that the requirements according to

ISO 9001:2015

are fulfilled. The certificate is valid from **2020-02-29** until **2023-02-27**.

Certificate Registration No.: **12 100 25208 TMS**.

Product Compliance Management
Munich, 2020-02-03





Product Service

Certificate

No. Q5 089752 0007 Rev. 00

Holder of Certificate: **Hartalega NGC Sdn. Bhd**
No.1, Persiaran Tanjung
Kawasan Perindustrian Tanjung
43900 Sepang, Selangor Darul Ehsan
MALAYSIA

Facility(ies): Hartalega NGC Sdn. Bhd
No.1, Persiaran Tanjung, Kawasan Perindustrian Tanjung, 43900
Sepang, Selangor Darul Ehsan, MALAYSIA

Certification Mark:



Scope of Certificate: Design and Development, Production and Distribution of
Natural Latex and Nitrile Powdered and Powder Free Non-
Sterile Examination Gloves

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: MYQMH0418038Rev1-721419970

Valid from: 2018-09-19
Valid until: 2020-11-26

Date, 2018-09-19

Stefan Preiß

510(K) Premarket Notification

1 to 10 of 82 Results

for Hartalega

1 2⁶ 3⁷ 4⁸ 5⁹ 6¹⁰ 7¹¹ 8¹² 9¹³ >¹⁴

10 results per page

[New Search](#)¹⁵

[Help](#)¹⁶

[Export To Excel](#) |

Device Name	▲ ¹⁷ ▼ ¹⁸	Applicant	▲ ¹⁹ ▼ ²⁰	510(K) Number ▲ ²¹ ▼ ²²	Decision Date ▲ ²³ ▼ ²⁴
Powdered Sterile Latex Surgical Glove, With Protein Content Labeling Claim (200 Micrograms Or Less)		HARTALEGA SDN BHD		K001959	07/26/2000
Powder Free Sterile Latex Surgical Gloves, Contains 50 Microgram Or Less Of Total Water Extractable Protein Per Gram		HARTALEGA SDN BHD		K002593	11/29/2000
Freeform Blue Powderfree Nitrile Examination Gloves		HARTALEGA SDN BHD		K022671	11/18/2002
Freeform Blue Powder-free Nitrile Examination Gloves		HARTALEGA SDN BHD		K041391	07/09/2004
Nitrile Powder Free Examination Gloves (White)		HARTALEGA SDN BHD		K050214	03/16/2005
Nitrile Powdered Examination Gloves (White)		HARTALEGA SDN BHD		K050215	03/11/2005
Chlorinated Powder Free Latex Examination Gloves (Yellow)		HARTALEGA SDN BHD		K050277	06/07/2005
Nitrile Powder Free Examination Gloves (Blue)		HARTALEGA SDN BHD		K051777	08/12/2005

Indications for Use

510(k) Number (if known)
K200581

Device Name
Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue)

Indications for Use (Describe)
Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue) is a non-sterile disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner. It is also tested to be used against Chemotherapy Drugs and Fentanyl Citrate.

These gloves were tested for use with chemotherapy drugs and Fentanyl Citrate as per ASTM D6978-05 (Reapproved 2013) Standard Practice for Assessment of Medical Gloves to Permeation by Chemotherapy Drugs.

Chemotherapy Drug and Concentration	Minimum Breakthrough Detection Time in Minutes
Carmustine (3.3 mg/ml)	21.4
Cisplatin (1.0 mg/ml)	>240
Cyclophosphamide (20.0 mg/ml)	>240
Dacarbazine (10.0 mg/ml)	>240
Doxorubicin Hydrochloride (2.0 mg/ml)	>240
Etoposide (20.0 mg/ml)	>240
Fluorouracil (50.0 mg/ml)	>240
Methotrexate (25.0 mg/ml)	>240
Mitomycin C (0.5 mg/ml)	>240
Paclitaxel (6.0 mg/ml)	>240
Thiotepa (10.0 mg/ml)	67.2
Vincristine Sulfate (1.0 mg/ml)	>240
Azacytidine (25.0 mg/ml)	>240
Carboplatin (10.0 mg/ml)	>240
Docetaxel (10 mg/ml)	>240
Epirubicin (2.0 mg/ml)	>240
Gemcitabine (38 mg/ml)	>240
Ifosfamide (50 mg/ml)	>240
Irinotecan (20 mg/ml)	>240
Mitoxantrone (2.0 mg/ml)	>240
Oncovin (1.0 mg/ml)	>240
Oxaliplatin (5 mg/ml)	>240
Vinorelbine (10 mg/ml)	>240

Please note that Carmustine and Thiotepa have extremely low permeation times of 21.4 minutes and 67.2 minutes respectively.

Warning: Do not use with Carmustine

Fentanyl Citrate and Concentration	Minimum Breakthrough Detection Time in Minutes
Fentanyl Citrate Injection (100 mcg/2ml)	>240

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Elizabeth F.
Claverie -S



CAPT Elizabeth Claverie, M.S.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure



April 25, 2020

Hartalega NGC SDN. BHD.
Nurul Kong
Senior Manager- Quality Assurance
Kawasan Perindustrian Tanjung
Sepang, Selangor 43900
Malaysia

Re: K200581

Trade/Device Name: Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with
Chemotherapy Drugs and Fentanyl Citrate (Blue)

Regulation Number: 21 CFR 880.6250

Regulation Name: Non-Powdered Patient Examination Glove

Regulatory Class: Class I, reserved

Product Code: LZA, LZC, QDO

Dated: February 27, 2020

Received: March 5, 2020

Dear Nurul Kong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

Indications for Use

510(k) Number (if known)
K200581

Device Name
Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue)

Indications for Use (Describe)
Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue) is a non-sterile disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner. It is also tested to be used against Chemotherapy Drugs and Fentanyl Citrate.

These gloves were tested for use with chemotherapy drugs and Fentanyl Citrate as per ASTM D6978-05 (Reapproved 2013) Standard Practice for Assessment of Medical Gloves to Permeation by Chemotherapy Drugs.

Chemotherapy Drug and Concentration	Minimum Breakthrough Detection Time in Minutes
Carmustine (3.3 mg/ml)	21.4
Cisplatin (1.0 mg/ml)	>240
Cyclophosphamide (20.0 mg/ml)	>240
Dacarbazine (10.0 mg/ml)	>240
Doxorubicin Hydrochloride (2.0 mg/ml)	>240
Etoposide (20.0 mg/ml)	>240
Fluorouracil (50.0 mg/ml)	>240
Methotrexate (25.0 mg/ml)	>240
Mitomycin C (0.5 mg/ml)	>240
Paclitaxel (6.0 mg/ml)	>240
Thiotepa (10.0 mg/ml)	67.2
Vincristine Sulfate (1.0 mg/ml)	>240
Azacytidine (25.0 mg/ml)	>240
Carboplatin (10.0 mg/ml)	>240
Docetaxel (10 mg/ml)	>240
Epirubicin (2.0 mg/ml)	>240
Gemcitabine (38 mg/ml)	>240
Ifosfamide (50 mg/ml)	>240
Irinotecan (20 mg/ml)	>240
Mitoxantrone (2.0 mg/ml)	>240
Oncovin (1.0 mg/ml)	>240
Oxaliplatin (5 mg/ml)	>240
Vinorelbine (10 mg/ml)	>240

Please note that Carmustine and Thiotepa have extremely low permeation times of 21.4 minutes and 67.2 minutes respectively.

Warning: Do not use with Carmustine

Fentanyl Citrate and Concentration
Fentanyl Citrate Injection (100 mcg/2ml)

Minimum Breakthrough Detection Time in Minutes
>240

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

510(k) Premarket Notification

[FDA Home](#) [Medical Devices](#) [Databases](#)



[510\(k\)](#) | [DeNovo](#) | [Registration & Listing](#) | [Adverse Events](#) | [Recalls](#) | [PMA](#) | [HDE](#) | [Classification](#) | [Standards](#)
[CFR Title 21](#) | [Radiation-Emitting Products](#) | [X-Ray Assembler](#) | [Medsun Reports](#) | [CLIA](#) | [TPLC](#)

[New Search](#)

[Back To Search Results](#)

Device Classification Name	Polymer Patient Examination Glove
510(K) Number	K133956
Device Name	NITRILE POWDER FREE EXAMINATION GLOVE WITH COLLOIDAL OATMEAL USP SKIN PROTECTANT DRUG - WHITE / DAWN BLUE / LEMON GREEN
Applicant	HARALEGA SDN BHD NO. 7, KAWASAN PERUSAHAAN SURIA Bestari Jaya, Selangor, MY 45600
Applicant Contact	Nurul Aisyah Kong
Correspondent	HARALEGA SDN BHD NO. 7, KAWASAN PERUSAHAAN SURIA Bestari Jaya, Selangor, MY 45600
Correspondent Contact	Nurul Aisyah Kong
Regulation Number	880.6250
Classification Product Code	LZA
Date Received	12/23/2013
Decision Date	05/28/2014
Decision	Substantially Equivalent (SESE)
Regulation Medical Specialty	General Hospital
510k Review Panel	General Hospital
Summary	Summary
Type	Traditional
Reviewed By Third Party	No
Combination Product	No

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Elizabeth F.
Claverie -S



CAPT Elizabeth Claverie, M.S.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Certificate of CE-Registration



This is to certify that, in accordance with the Medical Device Directive 93/42/EEC, Medical Device Safety Service GmbH (MDSS) agrees to perform all duties and responsibilities as the Authorized Representative for:

**Hartalega NGC Sdn. Bhd.
No. 1, Persiaran Tanjung
Kawasan Perindustrian Tanjung
43900 Sepang, Selangor
MALAYSIA**

as stipulated and demanded by the aforementioned Directive. The German Competent Authority has allocated the medical devices of the Manufacturer registration numbers as foreseen in:

Annex A dated January 18, 2019

The Manufacturer has provided MDSS with the appropriate Declaration(s) of Conformity confirming that the medical devices fulfill the applicable requirements of Directive 93/42/EEC. In compliance with German law, a safety officer has been appointed for Germany.

2019-01-18


Ludger Möller
President
MDSS GmbH

Testing. Development. Problem Solving.



April 15, 2009

• TEST REPORT •

PN 83672A - Amended

CHEMICAL ANALYTICAL SERVICES

Prepared For:
Hartalega SDN. BDH
Ms. Nurul Aisyah Kong
No. 7 Kawasan Perusahaan Suria
Bestari Jaya
Selangor, 45600
Malaysia

Prepared By


Tiffany L Heller
Chemical Technician

Approved By:


Ana C. Barbur, M.S.
Manager, Chemical & Pharmaceutical Services

An A2LA Accredited Testing Laboratory — Certificate Numbers 255.01 & 255.02
ISO 9001:2000 Registered
Member of ACIL: The American Council of Independent Laboratories



ISO 9001:2000
Registered



Letters and reports are for the exclusive use of the clients to whom they are addressed and shall not be reproduced, except in full, without the written permission of Akron Rubber Development Laboratory, Inc. (ARDL). The information contained herein applies to the specific material, products or processes tested or evaluated. No warranty of any kind is herein construed or implied. The liability of ARDL, Inc. shall be limited to the amount of consideration paid for services. ARDL, Inc. is accredited by A2LA for the test methods listed on the attached scope.

www.ardl.com

2887 Gilchrist Rd. | Akron, Ohio 44305 | answers@ardl.com
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

Testing, Development, Problem Solving.



April 15, 2009

Ms. Nurul Aisyah Kong
Hartalega SDN. BHD

Page 1 of 3 – PN 83672A - Amended

SUBJECT: Permeation testing per ASTM D 6978-05 on sample supplied by the above company. Wire Transfer.

RECEIVED: Glove sample identified as Nitrile Powder Free Examination Gloves (Blue) Code: ABLU

TESTING CHEMOTHERAPY DRUGS:

Table 1. List of the Testing Chemotherapy Drugs, Sources, and Expiration Dates

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carmustine (BCNU)	Sigma, Lot# 038K4008; Expiration 12/2009
Cisplatin	Sigma, Lot# 59H3657; Expiration 09/2009
Cyclophosphamide (Cytoxan)	Sigma, Lot# 058K1131; Expiration 1/2010
Dacarbazine (DTIC)	Hospira, Lot# U022223AA; Expiration 06/2010
Doxorubicin Hydrochloride	Teva, Lot#07N625; Expiration 10/2009
Etoposide (Toposar)	Teva, Lot# 31303976B; Expiration 9/2011
Fluorouracil	APP, Lot# 203867; Expiration 03/2010
Mitomycin C	Sigma, Lot# 048K1086; Expiration 01/2010
Methotrexate	Hospira, Lot# U024457AA; Expiration 05/2010
Paclitaxel (Taxol)	Dabur Oncology, Lot# PA08H00701; Exp. 05/2010
Thiotepa	Sigma, Lot#078K1526; Expiration 12/2009
Vincristine Sulfate	Hospira, Lot# U037139AA; Expiration 12/2009

COLLECTION MEDIA:

The collection media, which were selected, are listed in Table 2.

Table 2. Collection Media for Testing Chemotherapy Drugs

TEST DRUG AND CONCENTRATION	COLLECTION MEDIUM
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	10% Ethanol Aqueous Solution
Cisplatin, 1.0 mg/ml (1,000 ppm)	Distilled Water
Cyclophosphamide (Cytoxan), 20 mg/ml (20,000 ppm)	Distilled Water
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Distilled Water
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Distilled Water
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	Distilled Water
Fluorouracil, 50.0 mg/ml (50,000 ppm)	9.20 pH Sodium Hydroxide Solution
Methotrexate, 25 mg/ml (25,000 ppm)	Distilled Water
Mitomycin C, 0.5 mg/ml (500 ppm)	Distilled Water
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	30% Methanol Aqueous Solution
Thiotepa, 10.0 mg/ml (10,000 ppm)	Distilled Water
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	Distilled Water

www.ardl.com

2887 Glchrist Rd. | Akron, Ohio 44305 | answers@ardl.com
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

Annex A dated January 18, 2019
Manufacturer: Hartalega NGC Sdn. Bhd.

UMDNS Code Description Notified Medical Device Product Name & Catalogue Number	UMDNS Code	Risk Class	Cat. Code	Registration Number	NB No. / NB Certificate No.	NB Cert. Valid Until YYYY-MM-DD
Gloves, Examination/Treatment	11-882	I	10	DE/CA09/0170/H13/001-01	N.A.	N.A.
Latex Examination Gloves; Nitrile Examination Gloves; Antimicrobial Nitrile Powder Free Examination Gloves						
<i>Latex Powder Free Examination Gloves</i>						
<i>Latex Powdered Examination Gloves</i>						
<i>Nitrile Powder Free Examination Gloves</i>						
<i>Antimicrobial Nitrile Powder Free Examination Gloves</i>						



Testing, Development, Problem Solving.

April 15, 2009

Ms. Nurul Aisyah Kong
Hartalega SDN. BHD

Page 1 of 3 – PN 83672A - Amended

SUBJECT: Permeation testing per ASTM D 6978-05 on sample submitted by the above company. Wire Transfer.

RECEIVED: Glove sample identified as Nitrile Powder Free Examination Gloves (Blue) Code: ABLU.

TESTING CHEMOTHERAPY DRUGS:

Table 1. List of the Testing Chemotherapy Drugs, Sources, and Expiration Dates

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carmustine (BCNU)	Sigma, Lot# 038K4008, Expiration 12/2009
Cisplatin	Sigma, Lot# 59H3657, Expiration 09/2009
Cyclophosphamide (Cytoxan)	Sigma, Lot# 068K1131, Expiration 1/2010
Dacarbazine (DTIC)	Hospira, Lot# U022223AA, Expiration 06/2010
Doxorubicin Hydrochloride	Teva, Lot#07N625, Expiration 10/2009
Etoposide (Toposar)	Teva, Lot# 31303976B, Expiration 9/2011
Fluorouracil	APP, Lot# 203867, Expiration 03/2010
Mitomycin C	Sigma, Lot# 048K1088, Expiration 01/2010
Methotrexate	Hospira, Lot# U024457AA, Expiration 05/2010
Paclitaxel (Taxol)	Dabur Oncology, Lot# PA08H00701, Exp. 05/2010
Thiotepa	Sigma, Lot#078K1528, Expiration 12/2009
Vincristine Sulfate	Hospira, Lot# U037139AA, Expiration 12/2009

COLLECTION MEDIA:

The collection media, which were selected, are listed in Table 2.

Table 2. Collection Media for Testing Chemotherapy Drugs

TEST DRUG AND CONCENTRATION	COLLECTION MEDIUM
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	10% Ethanol Aqueous Solution
Cisplatin, 1.0 mg/ml (1,000 ppm)	Distilled Water
Cyclophosphamide (Cytoxan), 20 mg/ml (20,000 ppm)	Distilled Water
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Distilled Water
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Distilled Water
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	Distilled Water
Fluorouracil, 50.0 mg/ml (50,000 ppm)	9.20 pH Sodium Hydroxide Solution
Methotrexate, 25 mg/ml (25,000 ppm)	Distilled Water
Mitomycin C, 0.5 mg/ml (500 ppm)	Distilled Water
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	30% Methanol Aqueous Solution
Thiotepa, 10.0 mg/ml (10,000 ppm)	Distilled Water
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	Distilled Water

www.ardl.com

2887 Gilchrist Rd. | Akron, Ohio 44305 | answers@ardl.com
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

Testing. Development. Problem Solving.



April 15, 2009


• TEST REPORT •

PN 83672A - Amended

CHEMICAL ANALYTICAL SERVICES

Prepared For:
Hartalega SDN. BDH
Ms. Nurul Aisyah Kong
No. 7 Kawasan Perusahaan Suria
Bestari Jaya
Selangor, 45600
Malaysia

Prepared By


Tiffany L. Heller
Chemical Technician

Approved By:


Ana C. Barbur, M.S.
Manager, Chemical & Pharmaceutical Services

An A2LA Accredited Testing Laboratory — Certificate Numbers 255.01 & 255.02
ISO 9001:2000 Registered
Member of ACIL The American Council of Independent Laboratories



ISO 9001:2000
Registered



Letters and reports are for the exclusive use of the clients to whom they are addressed and shall not be reproduced, except in full, without the written permission of Akron Rubber Development Laboratory, Inc. (ARDL). The information contained herein applies to the specific material, products or processes tested or evaluated. No warranty of any kind is herein constituted or implied. The liability of ARDL, Inc. shall be limited to the amount of consideration paid for services. ARDL, Inc. is accredited by A2LA for the test methods listed on the attached scope.

www.ardl.com

2887 Gilchrist Rd. | Akron, Ohio 44305 | answers@ardl.com
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

Certifications

Gloveon's quality standards, management systems and exemplary regulatory compliance, all contribute to the global success of the company. Our capabilities have been assessed and certified by the following international governing bodies.



Management Service
ISO 9001:2015



America
ISO 13485:2016



EN ISO 13485:2016



Japan
Confirmation Letter
for GMP Audit



Product Service
EC Certificate



ISO 14001:2015



UL Certification



ISEGA Food Contact
Test
Certification (German)



Registration
Certificate
for Medical Device



NFPA Certification



510(k) Approval



PPE Cert



ANVISA

Testing. Development. Problem Solving.



April 15, 2009

• **TEST REPORT** •

PN 83672A - Amended

CHEMICAL ANALYTICAL SERVICES

Prepared For:
Hartalega SDN. BDH
Ms. Nurul Aisyah Kong
No. 7 Kawasan Perusahaan Suria
Bestari Jaya
Selangor, 45600
Malaysia

Prepared By


Tiffany L Heller
Chemical Technician

Approved By


Ana C. Barbur, M.S.
Manager, Chemical & Pharmaceutical Services

An A2LA Accredited Testing Laboratory — Certificate Numbers 255.01 & 255.02
ISO 9001:2000 Registered
Member of ACIL: The American Council of Independent Laboratories



ISO 9001:2000
Registered



Letters and reports are for the exclusive use of the clients to whom they are addressed and shall not be reproduced, except in full, without the written permission of Akron Rubber Development Laboratory, Inc. (ARDL). The information contained herein applies to the specific material, products or processes tested or evaluated. No warranty of any kind is herein construed or implied. The liability of ARDL, Inc. shall be limited to the amount of consideration paid for services. ARDL, Inc. is accredited by A2LA for the test methods listed on the attached scope.

www.ardl.com

2887 Gilchrist Rd. | Akron, Ohio 44305 | answers@ardl.com
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

This section applies only to requirements of the Paperwork Reduction Act of 1995.


DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.

FOR CUSTOMER OF TREASURE AND INFINITY CO.,LTD ONLY
DO NOT COPY



Certificate of CE-Registration

This is to certify that, in accordance with the Medical Device Directive 93/42/EEC, Medical Device Safety Service GmbH (MDSS) agrees to perform all duties and responsibilities as the Authorized Representative for:

**Hartalega NGC Sdn. Bhd.
No. 1, Persiaran Tanjung
Kawasan Perindustrian Tanjung
43900 Sepang, Selangor
MALAYSIA**

as stipulated and demanded by the aforementioned Directive. The German Competent Authority has allocated the medical devices of the Manufacturer registration numbers as foreseen in:

Annex A dated January 18, 2019

The Manufacturer has provided MDSS with the appropriate Declaration(s) of Conformity confirming that the medical devices fulfill the applicable requirements of Directive 93/42/EEC. In compliance with German law, a safety officer has been appointed for Germany.

2019-01-18


Ludger Möller
President
MDSS GmbH



Medical Device Safety Service

MDSS - Schiffgraben 41 - 30175 Hannover, Germany

Hartalega NGC Sdn. Bhd.
Khairunnisa Warsito
No. 1, Persiaran Tanjung
Kawasan Perindustrian Tanjung
43900 Sepang, Selangor
MALAYSIA

Schiffgraben 41
30175 Hannover, Germany
Tel: + 49 - 511 - 62 62 86 30
Fax: + 49 - 511 - 62 62 86 33
eMail: info@mdss.com
Internet: www.mdss.com

2019.01.18

Confirmation of CE Registration

Dear Khairunnisa,

It is our pleasure to enclose the new Certificate of CE-Registration for your product.

Please note that registration was performed under § 25 MPG (Medizinproduktegesetz). This is the *Federal Republic of Germany's national interpretation of Medical Device Directive 93/42/EEC*. Registration is therefore in accordance with EU legislation. We remind you that all products must meet the applicable provision of the European and national regulation before they may be placed on the market.

We are looking forward to continuing our good business relationship and wish you a successful product launch in Europe.

Best regards,

Juan Monferrer Tena
Administrative Assistant
Medical Device Safety Service GmbH

Encl.
1 Certificate of CE-Registration
1 Annex A

FOR CUSTOMER OF TREASURE AND WEALTHY CO., LTD ONLY
DO NOT COPY

MDSS - Medical Device Safety Service GmbH
Handelsregister Hannover HRB 57318 · USt-IdNr. DE 177346163 · Geschäftsführer: Ludger Möller
Bankverbindungen
Sparkasse Hannover
S.W.I.F.T.: SPHODE33
IBAN: DE24 2505 0180 0910 0792 77
Commerzbank AG, Hannover
S.W.I.F.T.: COBADE33
IBAN: DE67 2504 0066 0338 8816 00



MATERIAL SAFETY
DATA SHEET



SECTION 1: PRODUCT IDENTIFICATION

NAME Hartalega Sdn. Bhd.	ADDRESS C-G-9, Jalan Dataran SD1, Dataran SD, PJU 9, Bandar Sri Damansara, 52200 Kuala Lumpur
TELEPHONE NUMBER (603) 6277 1733	DATE PREPARED October 15, 2014
COMMON NAME (USED ON LABEL) Nitrile Powder Free Examination Gloves	CHEMICAL FAMILY Carboxylated Butadiene Acrylonitrile Polymer Latex
APPLICATION Medical and Dental	TRADENAME & SYNONYM GLOVEON COATS NITRILE (CTS38) NITRILE POWDER FREE EXAMINATION GLOVES COATS

SECTION 2: HAZARDOUS INGREDIENTS

HAZARDOUS COMPONENT	CAS #	%(WT)	TLV	PEL
N/A	N/A	N/A	N/A	N/A

PEL: Permissible Exposure Limit established by Occupational Safety and Health Administration (OSHA).
TLV: Threshold Limit Value established by the American Conference of Governmental Industrial Hygienists, 1987-1988.

SECTION 3: COMPOSITION/ INFORMATION ON INGREDIENTS

CHEMICAL COMPOSITION
All chemicals used are non-toxic/ non-hazardous.
Butadiene-Acrylonitrile Latex, Sodium Dodecylbenzenesulfonate, Sulphur, Zinc Oxide, Zinc Di-n-butylidithiocarbamate, Titanium Dioxide, Paraffin Wax Emulsion
Coating Ingredient
Colloidal Oatmeal & Constituents, Sodium Benzoate, Processing Aid

SECTION 4: FIRST AID MEASURE

If reaction in the form of skin irritation is noticed, remove gloves immediately and wash affected part with saline water. If there is no relief, seek medical reactions.

SECTION 5: FIRE FIGHTING MEASURE

FLASHPOINT	AUTOIGNITION TEMPERATURE	FLAMMABLE LIMITS IN AIR
N/A	N/A	N/A

EXTINGUISHING MEDIA
Chemical foam and dry chemical may be used.

FIRE-FIGHTING PROCEDURES
Use standard procedures for combustion material fires, including approved self-contained breathing apparatus.

FIRE AND EXPLOSION HAZARDS
No fire or explosion hazards are associated with these products. They will melt at elevated temperatures.

SECTION 6: ACCIDENTAL RELEASE MEASURES

BIOCOMPATABILITY
The chemical formulation of the gloves and surface lubrication materials does not contain any substances normally known to be harmful to the user or to any person with whom the gloves come into contact.

MEDICAL CONDITIONS GENERALLY AGGRAVATED BY EXPOSURE
Nitrile Powder Free Gloves are not expected to cause any adverse health effects.

SECTION 7: HANDLING AND STORAGE

PRECAUTIONS TO BE TAKEN IN HANDLING AND STORAGE
Store in a dry, cool and ventilated area. Do not store above 104 °F (40 °C). Shield open box from direct sunlight, fluorescent lighting and x-rays. Improper storage will decrease usable life.

SECTION 8: EXPOSURE CONTROLS/ PERSONAL PROTECTION					
EYE PROTECTION Not necessary under conditions of intended use.			SKIN PROTECTION Not necessary under conditions of intended use.		
RESPIRATORY PROTECTION Not necessary under conditions of intended use.			VENTILATION Not necessary under conditions of intended use.		
STEPS TO BE TAKEN IN CASE MATERIAL IS LEAKED OR SPILLED These products are solid articles and are not subject to leaks or spills.					
SECTION 9: PHYSICAL AND CHEMICAL PROPERTIES					
APPEARANCE/ ODOR Ambidextrous, Beaded Cuff, Micro-textured, Chlorinated, Powder Free, Coated with Colloidal Oatmeal USP Skin Protectant, Dawn Blue.					
DIMENSION	X-SMALL	SMALL	MEDIUM	LARGE	X-LARGE
Length (mm)	Minimum 230 (same for all)				
Width (mm)	76 ± 4	86 ± 4	98 ± 4	107 ± 4	115 ± 4
THICKNESS (mm) - SINGLE WALL MEASUREMENT (same for all)					
Finger (mm)	0.09 ± 0.02				
Palm (mm)	0.07 ± 0.02				
TENSILE PROPERTIES		UNAGED		AGED	
Tensile Strength (Mpa)		Min. 18.0 MPa		Min. 16.0 MPa	
Ultimate Elongation (%)		Min. 500%		Min. 400%	
SECTION 10: STABILITY AND REACTIVITY					
BOILING POINT N/A		VAPOR PRESSURE (mm Hg) N/A		VAPOR DENSITY (air=1) N/A	
SPECIFIC GRAVITY (water=1) N/A		SOLUBILITY IN WATER Insoluble		% VOLATILE BY VOLUME N/A	
EVAPORATION RATE N/A			VISCOSITY N/A		
SECTION 11: TOXICOLOGICAL INFORMATION					
STABILITY Stable.			CONDITIONS TO AVOID Does not apply.		
INCOMPATIBILITY (MATERIALS TO AVOID) High polar solvent like methyl ethyl ketone, acetone.					
HAZARDOUS DECOMPOSITION PRODUCTS In a fire, these products may produce a black smoke. Carbon Dioxide, Carbon Monoxide, Oxides of Nitrogen, aromatic/aliphatic hydrocarbons.					
HAZARDOUS POLYMERIZATION Will not occur.					
SECTION 12: ECOLOGICAL INFORMATION					
N/A					
SECTION 13: DISPOSAL CONSIDERATION					
WASTE DISPOSAL METHOD Consult current local, state and federal regulations for proper disposal methods.					
SECTION 14: TRANSPORT INFORMATION					
N/A					
SECTION 15: REGULATORY INFORMATION					
N/A					
SECTION 16: OTHER INFORMATION					
RECOMMENDED USE AND RESTRICTION The Nitrile Powder Free Gloves is a Single Use device.					